

Sample Calibration Report

Report number: Sample

Model: Portable One
Serial Number: Sample

Data Type: AS RECEIVED
Date of Cal: 28-Feb-2008

Program:
 P1-ATS1Cal-apc.apb 1.00

Option Status

IMD	Not Installed
EURZ	Not Installed
PDX	Installed
P1D	Installed

Sample

Explanatory notes to the last three columns of the calibration report

"MU" - The column labeled "MU" lists the expanded measurement uncertainties derived from published specifications of the calibration equipment, resolution factors, and statistical data of repeatability which may include inter-lab comparisons and/or other information. These are based on a coverage factor of 2 (k=2) corresponding to a confidence level of 95%.

"TUR" - The column labeled "TUR" lists the test uncertainty ratio calculated by dividing the lesser of the lower and upper reading tolerances by the expanded measurement uncertainty. An entry of "na" indicates [1] the specified limits are one-sided, or [2] the performance characteristic is a self-test or system specification that is not independently certifiable.

"Result" - The column labeled "Result" lists color-coded assessments that the observed characteristic is within its specified limits of performance. There are three possible indications:

pass -- The *READING* is within the specified upper and lower limits reduced by guard-bands equal to the expanded measurement uncertainty or "MU". The confidence level is 95% or higher the observed characteristic is within specification.

pass (<95%) -- The *READING* is within the specified upper and lower limits, but it is close to one of the limits by an amount that is less than the expanded measurement uncertainty or "MU". Conformance to specification is more probable than not, however the confidence level may be <95%.

>> FAIL << -- The observed characteristic is out of specification.

This report is valid only when accompanied by a signed Certificate of Calibration.

File name: P1-ATS1Report-apc.xls

File version: REV1.00-18FEB2008-SR

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
ANALOG GENERATOR CHARACTERISTICS							
[1] Sine Frequency Accuracy (Hz, kHz)							
	10 Hz	9.950	10.005	10.050	0.005	>10	pass
	100 Hz	99.50	99.97	100.50	0.05	>10	pass
	1.00 kHz	0.9950	0.9997	1.0050	0.0005	>10	pass
	10.0 kHz	9.950	9.998	10.050	0.005	>10	pass
	100 kHz	99.50	99.98	100.50	0.05	>10	pass
[2] Sine Amplitude Accuracy at 1 kHz (Volts, mVolts)							
Channel A	10.0 V	9.7700	10.0166	10.2300	0.0061	>10	pass
	2.50 V	2.4425	2.5016	2.5575	0.0015	>10	pass
	1.00 V	0.9770	1.0005	1.0230	0.0020	>10	pass
	250 mV	244.25	250.73	255.75	0.50	>10	pass
	100 mV	97.70	100.25	102.30	0.20	>10	pass
	25 mV	24.425	25.081	25.575	0.050	>10	pass
Channel B	10.0 V	9.7700	10.0165	10.2300	0.0061	>10	pass
	2.50 V	2.4425	2.5016	2.5575	0.0015	>10	pass
	1.00 V	0.9770	1.0005	1.0230	0.0020	>10	pass
	250 mV	244.25	250.73	255.75	0.50	>10	pass
	100 mV	97.70	100.25	102.30	0.20	>10	pass
	25 mV	24.425	25.081	25.575	0.050	>10	pass
[3] Sine Flatness at 2.5 Volts (dB)							
Channel A	10 Hz	-0.0500	0.0169	0.0500	0.0026	>10	pass
	20 Hz	-0.0500	0.0052	0.0500	0.0026	>10	pass
	20 kHz	-0.0500	-0.0008	0.0500	0.0027	>10	pass
	50 kHz	-0.3000	-0.0132	0.3000	0.0037	>10	pass
	120 kHz	-0.300	-0.017	0.300	0.028	>10	pass
[4] Sine THD+N (dB) - self test							
26.25V, 80k BW	25 Hz	-999	-98.7	-92.0	2.0	na	pass
	1 kHz	-999	-98.5	-92.0	2.0	na	pass
	20 kHz	-999	-96.3	-92.0	2.0	na	pass
1.00V, Full BW	10 Hz	-999	-99.4	-80.0	2.0	na	pass
	50 kHz	-999	-90.4	-80.0	2.0	na	pass
[5] Squarewave Amplitude Accuracy (Volts)							
	1.00 V, 100 Hz	1.3647	1.4232	1.4637	0.0024	>10	pass
[6] Source Resistance Accuracy (Ohms)							
Channel A	40 Ohm Unbal	38.00	41.03	42.00	0.020	>10	pass
	40 Ohm Bal	38.00	41.15	42.00	0.020	>10	pass
	150 Ohm Bal	148.00	151.04	152.00	0.040	>10	pass
	600 Ohm Bal	598.00	601.13	602.00	0.12	>10	pass
Channel B	40 Ohm Unbal	38.00	40.91	42.00	0.020	>10	pass
	40 Ohm Bal	38.00	41.03	42.00	0.020	>10	pass
	150 Ohm Bal	148.00	150.93	152.00	0.040	>10	pass
	600 Ohm Bal	598.00	601.24	602.00	0.12	>10	pass
ANALOG ANALYZER CHARACTERISTICS							
[7] Input Termination Accuracy (Ohms)							
Channel A	600 Ohms	594.00	603.15	606.00	0.12	>10	pass
Channel B	600 Ohms	594.00	603.74	606.00	0.12	>10	pass
[8] Input Common Mode Rejection (dB)							
Channel A	2.5V range, 50 Hz	-999	-103.5	-70.0	2.0	na	pass
	2.5V range, 1 kHz	-999	-95.4	-70.0	2.0	na	pass
	2.5V range, 20 kHz	-999	-72.1	-70.0	2.0	na	pass
	10V range, 1kHz	-999	-75.4	-50.0	2.0	na	pass
	25V range, 1 kHz	-999	-75.4	-50.0	2.0	na	pass
	100V range, 1kHz	-999	-74.8	-50.0	2.0	na	pass
	250V range, 1kHz	-999	-67.2	-50.0	2.0	na	pass
	Channel B	2.5V range, 50 Hz	-999	-73.7	-70.0	2.0	na
2.5V range, 1 kHz		-999	-73.7	-70.0	2.0	na	pass
2.5V range, 20 kHz		-999	-72.7	-70.0	2.0	na	pass
10V range, 1kHz		-999	-63.7	-50.0	2.0	na	pass
25V range, 1 kHz		-999	-63.2	-50.0	2.0	na	pass
100V range, 1kHz		-999	-64.2	-50.0	2.0	na	pass
250V range, 1kHz		-999	-63.3	-50.0	2.0	na	pass

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
ANALOG ANALYZER, Continued							
[9] Level/Amplitude Meter Accuracy at 1 kHz (Volts, mVolts)							
Channel A	140 V	138.4	139.7	141.6	0.2	8.0	pass
	60 V	59.31	59.87	60.69	0.06	>10	pass
	20 V	19.77	19.96	20.23	0.02	>10	pass
	6 V	5.931	5.984	6.069	0.006	>10	pass
	2 V	1.977	1.996	2.023	0.002	>10	pass
	600 mV	593.0	598.7	607.0	0.6	>10	pass
	200 mV	197.6	199.6	202.4	0.2	>10	pass
	60 mV	59.21	59.69	60.79	0.06	>10	pass
Channel B	140 V	138.4	139.9	141.6	0.2	8.0	pass
	60 V	59.31	59.93	60.69	0.06	>10	pass
	20 V	19.77	19.98	20.23	0.02	>10	pass
	6 V	5.931	5.989	6.069	0.006	>10	pass
	2 V	1.977	1.998	2.023	0.002	>10	pass
	600 mV	593.0	599.1	607.0	0.6	>10	pass
	200 mV	197.6	199.6	202.4	0.2	>10	pass
	60 mV	59.21	59.83	60.79	0.06	>10	pass
10 mV	9.79	9.97	10.22	0.01	>10	pass	
[10] Level/Amplitude Meter Flatness at 2 Volts (dB)							
Channel A	10 Hz	-0.200	-0.009	0.200	0.020	10	pass
	20 Hz	-0.050	0.004	0.050	0.010	5.0	pass
	20 kHz	-0.050	0.000	0.050	0.010	5.0	pass
	50 kHz	-0.200	-0.035	0.200	0.010	>10	pass
	120 kHz	-0.500	-0.283	0.500	0.027	>10	pass
Channel B	10 Hz	-0.200	-0.048	0.200	0.020	10	pass
	20 Hz	-0.050	-0.009	0.050	0.010	5.0	pass
	20 kHz	-0.050	0.000	0.050	0.010	5.0	pass
	50 kHz	-0.200	0.009	0.200	0.010	>10	pass
	120 kHz	-0.500	0.052	0.500	0.027	>10	pass
[11] Bandwidth Limiting Filters (dB)							
22 Hz - 22 kHz	22.4 Hz	-6.00	-1.32	0.00	0.04	na	pass
	31.5 Hz	-0.50	-0.06	0.50	0.04	>10	pass
	16 kHz	-0.50	-0.19	0.50	0.04	>10	pass
	22.4 kHz	-6.00	-4.04	0.00	0.04	na	pass
400 Hz Highpass	400 Hz	-3.70	-3.18	-2.40	0.04	na	pass
30 kHz Lowpass	30 kHz	-3.70	-3.15	-2.40	0.04	na	pass
80 kHz Lowpass	80 kHz	-3.70	-3.33	-2.40	0.04	na	pass
[12] Noise Weighting Filters (dB)							
A-Weighting	100 Hz	-20.10	-19.27	-18.10	0.04	na	pass
	1 kHz	-0.20	-0.08	0.20	0.02	10	pass
	10 kHz	-6.50	-2.56	-0.50	0.04	na	pass
CCIR-RMS	100 Hz	-20.80	-19.84	-18.80	0.04	na	pass
	1 kHz	-0.20	0.01	0.20	0.02	10	pass
	6.3 kHz	11.70	12.19	12.70	0.04	na	pass
	12.5 kHz	-1.20	-0.17	1.20	0.04	>10	pass
CCIR-QPK	1 kHz	-0.20	0.09	0.20	0.02	10	pass
CCIR-ARM	2 kHz	-0.20	-0.04	0.20	0.02	10	pass
[13] Residual Noise (uVolts) - self test							
Channel A	22-22k BW	0	1.01	1.50	0.06	na	pass
	A-weighted	0	0.72	1.00	0.06	na	pass
	CCIR-qpk	0	3.21	5.00	0.20	na	pass
Channel B	22-22k BW	0	1.01	1.50	0.06	na	pass
	A-weighted	0	0.72	1.00	0.06	na	pass
	CCIR-qpk	0	3.19	5.00	0.20	na	pass
[14] Selective Amplitude Accuracy at 1.00 Vrms (dBV)							
	20 Hz	-0.30	-0.08	0.30	0.02	>10	pass
	1 kHz	-0.30	0.00	0.30	0.02	>10	pass
	20 kHz	-0.30	-0.04	0.30	0.02	>10	pass
	120 kHz	-0.30	-0.16	0.30	0.05	6.0	pass

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
ANALOG ANALYZER, Continued							
[15] THD+N/SINAD Accuracy (dBV)							
20 Hz notch	40 Hz	-1.00	-0.39	1.00	0.02	>10	pass
	60 Hz	-1.00	-0.14	1.00	0.02	>10	pass
	1 kHz	-1.00	-0.03	1.00	0.02	>10	pass
	20 kHz	-1.00	-0.02	1.00	0.02	>10	pass
1 kHz notch	20 Hz	-1.00	-0.10	1.00	0.02	>10	pass
	2.0 kHz	-1.00	-0.35	1.00	0.02	>10	pass
	3.0 kHz	-1.00	-0.12	1.00	0.02	>10	pass
20 kHz notch	20 kHz	-1.00	-0.03	1.00	0.02	>10	pass
	1 kHz	-1.00	-0.03	1.00	0.02	>10	pass
	40 kHz	-1.00	-0.39	1.00	0.02	>10	pass
	60 kHz	-1.00	-0.18	1.00	0.02	>10	pass
	120 kHz	1.00	0.28	1.00	0.05	>10	pass
[16] - Phase Measurement Offset (Degrees)							
	10 Hz	-5.0	-0.1	5.0	0.6	8.3	pass
	20 Hz	-2.0	-0.1	2.0	0.2	10	pass
	1 kHz	-2.0	-0.1	2.0	0.2	10	pass
	20 kHz	-2.0	-0.5	2.0	0.2	10	pass
	50 kHz	-5.0	0.0	5.0	0.2	>10	pass
[17] Frequency Measurement Accuracy (µHz/Hz)							
	9.900 kHz	-100.0	-36.1	100.0	10.0	10	pass
[18] Input Residual Crosstalk at 20 kHz (dB)							
Ch B into Ch A		-999	-132.3	-120.0	6.0	na	pass
Ch A into Ch B		-999	-148.7	-120.0	6.0	na	pass
[19] Residual Wow & Flutter (%) - self-test							
IEC	Weighted	0%	0.000%	0.005%	0.001%	na	pass
	Unweighted	0%	0.000%	0.010%	0.001%	na	pass
OPTION "IMD" RELATED				>> Option Not Installed			
[20] Generator IMD Amplitude Accuracy at 1.000 Vrms (Volts)							
SMPTE 4:1	70Hz/8kHz						
[21] Residual SMPTE IMD (%) - self test							
Channel A	25.25 V						
	1.00 V						
Channel B	25.25 V						
	1.00 V						
AES/EBU DIGITAL I/O CHARACTERISTICS							
[22] Output Voltage Accuracy (Volts, mVolts) - oscilloscope referenced							
Balanced	4.00 Vpp	3.520	4.016	4.480	0.050	9.6	pass
	0.40 Vpp	0.280	0.414	0.520	0.005	>10	pass
Unbalanced	1.50 Vpp	1.330	1.420	1.670	0.022	6.4	pass
	0.20 Vpp	0.160	0.192	0.240	0.003	>10	pass
[23] Input Voltage Measurement Accuracy (Volts) - oscilloscope referenced							
Balanced	4.00 Vpp	3.55	4.06	4.45	0.06	4.2	pass
	0.40 Vpp	0.31	0.43	0.49	0.01	>10	pass
Unbalanced	1.50 Vpp	1.32	1.53	1.68	0.02	4.4	pass
	0.20 Vpp	0.15	0.21	0.25	0.01	>10	pass
[24] Jitter Accuracy at 500Hz, 48ks (UI) - self test							
	2.450 UIpk	2.165	2.499	2.735	0.040	na	pass
	0.510 UIpk	0.419	0.505	0.601	0.010	na	pass
	0.170 UIpk	0.113	0.174	0.227	0.004	na	pass
[25] Jitter Flatness at 0.3 UI, 48 ks (dB) - self test							
	100 Hz	-1.50	0.89	1.50	0.10	na	pass
	200 Hz	-1.50	0.47	1.50	0.10	na	pass
	2.5 kHz	-1.50	0.09	1.50	0.10	na	pass
	5 kHz	-1.50	0.12	1.50	0.10	na	pass
	10 kHz	-1.50	0.17	1.50	0.10	na	pass
	22 kHz	-1.50	0.06	1.50	0.14	na	pass
[26] Residual Jitter, 700 Hz to 30 kHz BW, 48 ks (nsec) - self test							
	RMS	0	0.17	0.81	0.10	na	pass
	PEAK	0	0.44	2.43	0.20	na	pass

END OF REPORT